



Quality Control Certificate

Etanercept ELISA (mAb-based) AB102

| | | | |
|---------------------|----------------------|--|--|
| REF IG-AB102 | LOT AB102-102 |  16.05.2024 | IVD  |
|---------------------|----------------------|--|--|

| Standard Curve | | | |
|----------------|--|--------------------------------|----------|
| 10X | 1:10 Diluted Study Concentration (ng/mL) | Measured Value OD 450/620nm | Range OD |
| Standard A | 200 | 2,795 | ≥ 1.500 |
| Standard B | 60 | 1,192 | - |
| Standard C | 20 | 0,431 | - |
| Standard D | 6 | 0,157 | - |
| Standard E | 0 | 0,048 | ≤ 0.100 |

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

QC: PASS

Validation Date: 13.05.2022

