



Quality Control Certificate

Tocilizumab ELISA (mAb-based) AB108

REF IG-AB108	LOT AB108-105	 20.10.2025	IVD 
---------------------	----------------------	--	--

Standard Curve			
10X	1:10 Diluted Study Concentration (ng/mL)	Measured Value OD 450/620nm	Range OD
Standard A	200	3.170	≥ 1.500
Standard B	60	1.266	-
Standard C	20	0.479	-
Standard D	6	0.189	-
Standard E	0	0.065	≤ 0.100

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date (dd-mm-yy): 20.10.2023

QC: PASS

