


## Quality Control Certificate

### Ranibizumab ELISA (mAb-based) AB110a

|                      |                       |   |  |
|----------------------|-----------------------|---|--|
| <b>REF</b> IG-AB110a | <b>LOT</b> AB110a-106 |  27.07.2025 (dd,mm,yyyy) |  |
|----------------------|-----------------------|---|--|

| Standard Curve |  |                                |              |
|----------------|--|--------------------------------|--------------|
| 10X            | 1:10 Diluted<br>Study Concentration<br>(ng/mL) | Measured Value<br>OD 450/620nm | Range OD     |
| Standard A     | 30   | 3.536                          | $\geq 1.500$ |
| Standard B     | 10   | 1.914                          | -            |
| Standard C     | 3  | 0.674                          | -            |
| Standard D     | 1  | 0.231                          | -            |
| Standard E     | 0  | 0.042                          | $\leq 0.100$ |

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date (dd,mm,yyyy): 27.07.2023

**QC: PASS**

