



Quality Control Certificate

Ustekinumab ELISA (mAb-based) AB121

REF IG-AB121	LOT AB121-204	 01.03.2024	IVD 
---------------------	----------------------	--	--

Standard Curve			
10X	1:10 Diluted Study Concentration (ng/mL)	Measured Value OD 450/620nm	Range OD
Standard A	100	2.301	≥ 1.500
Standard B	30	0.958	-
Standard C	10	0.339	-
Standard D	3	0.126	-
Standard E	0	0.038	≤ 0.100

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date: 01.03.2022

QC: PASS


AybayTech Biyoteknoloji
İmalat San. ve Tic. Ltd. Şti.
Meydanlar Bulvarı ATB İş Merkezi
No:1/285 Çankaya/ANKARA
Ostim V.D. 109 099 3570
Mersis No:0109099357000001
www.aybaytech.com