



Certificate of Analysis (CoA)

Antibody to Infliximab ELISA BB101

| | | | |
|---------------------|----------------------|---|--|
| REF IG-BB101 | LOT BB101-209 |  10.01.2025 (dd,mm,yy) | IVD  |
|---------------------|----------------------|---|--|

| Standards/Positive Control/Cut-Off | | | |
|------------------------------------|---------------------------------------|-----------------------------|-------------|
| | Ready-to-use Concentration (ng/mL) | Measured Value (OD450nm) | Range OD |
| Standard A | 30 | 1.812 | >0,500 |
| Standard B | 10 | 0.707 | >Standard C |
| Standard C | 3 | 0.254 | >Standard D |
| Standard D | 1 | 0.164 | >Standard E |
| Standard E | 0 | 0.044 | <0,100 |
| Calibrator | | | <0,100 |
| Positive Control | 16.5 ng/mL | 1.046 | ± 20% |

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date: 10.01.2023

QC: PASS

