

# Certificate of Analysis (CoA)

## Antibody to Infliximab ELISA BB101

<b>REF</b> IG-BB101	<b>LOT</b> BB101-210	 07.11.2025 (dd,mm,yy)	<b>IVD</b> 
---------------------	----------------------	---	--

Standards/Positive Control/Cut-Off			
	Ready-to-use Concentration (ng/mL)	Measured Value (OD450nm)	Range OD
Standard A	30	1.481	>0,500
Standard B	10	0.582	>Standard C
Standard C	3	0.222	>Standard D
Standard D	1	0.157	>Standard E
Standard E	0	0.061	<0,100
Calibrator			<0,100
Positive Control	16.52 ng/mL	0.860	± 20%

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date: 07.11.2023

**QC: PASS**

