



Quality Control Certificate

Antibody to Etanercept ELISA BB102

REF IG-BB102	LOT BB102-203	 06.09.2024	IVD 
---------------------	----------------------	--	--

Standards/Positive Control/Cut-Off			
	Ready-to-use Concentration (ng/mL)	Measured Value (OD450nm)	Range OD
Standard A	1000	3.235	>0,500
Standard B	300	1.323	>Standard C
Standard C	100	0.667	>Standard D
Standard D	30	0.202	>Standard E
Standard E	0	0.055	<0,100
Positive Control	551 ng/mL ($\pm 20\%$)	1.922	
Cut-Off (CO)			StE<CO<StD

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date:06.09.2022

QC: PASS

