



## Quality Control Certificate

### Antibody to Bevacizumab ELISA BB104

<b>REF</b>	IG-BB104	<b>LOT</b>	BB104-203		06.09.2024	<b>IVD</b>	
------------	----------	------------	-----------	---	------------	------------	---

Standards/Positive Control/Cut-Off			
	Ready-to-use Concentration (ng/mL)	Measured Value (OD450nm)	Range OD
Standard A	1000	1.325	>0,500
Standard B	300	0.823	>Standard C
Standard C	100	0.364	>Standard D
Standard D	30	0.181	>Standard E
Standard E	0	0.057	<0,100
Positive Control	283 ng/mL ( $\pm 20\%$ )	0.548	
Cut-Off (CO)			StE<CO<StD

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date:06.09.2022

**QC: PASS**

