



Quality Control Certificate

Antibody to Golimumab ELISA BB107

REF IG-BB107	LOT BB107-204	 05.05.2024	IVD 
---------------------	----------------------	--	--

Standards/Positive Control/Cut-Off			
	Ready-to-use Concentration (ng/mL)	Measured Value (OD450nm)	Range OD
Standard A	40	1,141	>0,500
Standard B	20	0,545	>Standard C
Standard C	10	0,332	>Standard D
Standard D	5	0,185	>Standard E
Standard E	0	0,048	<0,100
Positive Control	20,8 ng/mL (\pm 20%)	0,606	
Cut-Off (CO)			StE<CO<StD

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date:05.05.2022

QC: PASS


 AybayTech Biyoteknoloji
 İmalat San. ve Tic. Ltd. Şti.
 Mecidiyeköy Mah. Balı Bulvarı ATB İş Merkezi
 No:1/285 Kat:11/11/11/ANKARA
 Ostim V.D. 109 099 3570
 Mersis No:0109099357000001
 www.aybaytech.com