



## Certificate of Analysis (CoA)

### Antibody to Certolizumab Pegol ELISA BB109

<b>REF</b>	IG-BB109	<b>LOT</b>	BB109-204		28.10.2024	<b>IVD</b>	
------------	----------	------------	-----------	---	------------	------------	---

Standards/Positive Control/Cut-Off			
	Ready-to-use Concentration (ng/mL)	Measured Value (OD450nm)	Range OD
Standard A	1000	1.245	>0,500
Standard B	300	0.592	>Standard C
Standard C	100	0.223	>Standard D
Standard D	30	0.104	>Standard E
Standard E	0	0.047	<0,100
Calibrator		0.049	<0,100
Positive Control	227 ng/mL	0.378	± 20%

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date:28.10.2022

**QC: PASS**

