



Certificate of Analysis

Antibody to Vedolizumab ELISA BB116

REF IG-BB116	LOT BB116-209	 03.01.2025	IVD 
---------------------	----------------------	--	--

Standards/Positive Control/Cut-Off			
	Ready-to-use Concentration (ng/mL)	Measured Value (OD450nm)	Range OD
Standard A	150	1.772	>0,500
Standard B	50	0.678	>Standard C
Standard C	15	0.235	>Standard D
Standard D	5	0.129	>Standard E
Standard E	0	0.044	<0,100
Calibrator			<0,100
Positive Control	41.5 ng/mL	0.542	± 20%

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date:03.01.2023

QC: PASS

