



Certificate of Analysis

Antibody to Vedolizumab ELISA BB116

REF IG-BB116	LOT BB116-210	 02.02.2025	IVD 
---------------------	----------------------	--	--

Standards/Positive Control/Cut-Off			
	Ready-to-use Concentration (ng/mL)	Measured Value (OD450nm)	Range OD
Standard A	150	1.313	>0,500
Standard B	50	0.512	>Standard C
Standard C	15	0.191	>Standard D
Standard D	5	0.112	>Standard E
Standard E	0	0.050	<0,100
Calibrator			<0,100
Positive Control	41 ng/mL	0.416	± 20%

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date:02.02.2023

QC: PASS

