



## **Quality Control Certificate**

## Rituximab ELISA (mAb-based) AB106

REF IG-AB106	LOT AB106-105	11.11.2023	IVD (€
--------------	---------------	------------	--------

Standard Curve				
10X	1:10 Diluted	Measured Value	Range OD	
	Study Concentration (ng/mL)	OD 450/620nm		
Standard A	200	2.925	≥ 1.500	
Standard B	60	1.121	-	
Standard C	20	0.429	-	
Standard D	6	0.166	-	
Standard E	0	0.043	≤ 0.100	

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date: 11.11.2021

QC: PASS

MybayTech Biyoteknoloji malat San. ve Tic. Ltd. Sti. burble Ball Bulvari ATB İş Merkezi Nc.1/285 Teamar alle/ANKARA Ostim V.D. 109 099 3570 Mersis No:0109099357000001 www.aybaytech.com