



Quality Control Certificate

Rituximab ELISA (mAb-based) AB106

REF IG-AB106	LOT AB106-106	 10.08.2024	IVD 
---------------------	----------------------	--	--

Standard Curve			
10X	1:10 Diluted Study Concentration (ng/mL)	Measured Value OD 450/620nm	Range OD
Standard A	200	2.699	≥ 1.500
Standard B	60	1.247	-
Standard C	20	0.534	-
Standard D	6	0.178	-
Standard E	0	0.056	≤ 0.100

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date: 10.08.2022

QC: PASS

