



Certificate of Analysis (CoA)

Antibody to Nivolumab ELISA BB113

REF IG-BB113	LOT BB113-202	 20.02.2026 (dd,mm,yy)	IVD 
---------------------	----------------------	---	--

Standards/Positive Control/Cut-Off			
	Ready-to-use Concentration (ng/mL)	Measured Value (OD450nm)	Range OD
Standard A	300	1.748	>0,500
Standard B	100	0.688	>Standard C
Standard C	30	0.269	>Standard D
Standard D	10	0.165	>Standard E
Standard E	0	0.07	<0,150
Calibrator			<0,150
Positive Control	167 ng/mL	1.023	± 20%

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date: 20.02.2024

QC: PASS

