



Quality Control Certificate

Infliximab ELISA (mAb-based) AB101

| | | | |
|---------------------|----------------------|--|--|
| REF IG-AB101 | LOT AB101-209 |  15.06.2026 | IVD  |
|---------------------|----------------------|--|--|

| Standard Curve | | | |
|----------------|--|--------------------------------|----------|
| 10X | 1:10 Diluted Study Concentration (ng/mL) | Measured Value OD 450/620nm | Range OD |
| Standard A | 150 | 2.470 | ≥ 1.500 |
| Standard B | 50 | 1.072 | - |
| Standard C | 15 | 0.359 | - |
| Standard D | 5 | 0.152 | - |
| Standard E | 0 | 0.055 | ≤ 0.100 |

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date (dd,mm,yyyy): 15.06.2024

QC: PASS

