



Certificate of Analysis (CoA)

Aflibercept ELISA AA115

REF IG-AA115	LOT AA115-124	 10.07.2026 (dd.mm.yy)	IVD 
---------------------	----------------------	---	--

Standard Curve			
10X	1:10 Diluted Study Concentration (ng/mL)	Measured Value OD 450/620nm	Range OD
Standard A	200	3.553	≥ 1.500
Standard B	60	1.046	-
Standard C	20	0.246	-
Standard D	6	0.070	-
Standard E	0	0.015	≤ 0.100

The test results are valid, as long as all criterias are met.

1

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date: 10.07.2024

QC: PASS

