

Certificate of Analysis (CoA)

Aflibercept ELISA AA115

REF IG-AA115	LOT AA115-128	 05.05.2027 (dd.mm.yy)	IVD 
---------------------	----------------------	---	--

Standard Curve			
10X	1:10 Diluted Study Concentration (ng/mL)	Measured Value OD 450/620nm	Range OD
Standard A	200	3.200	≥ 1.500
Standard B	60	1.107	-
Standard C	20	0.278	-
Standard D	6	0.081	-
Standard E	0	0.019	≤ 0.100

The test results are valid, as long as all criterias are met.
1

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date: 05.05.2025

QC: PASS

