



Quality Control Certificate

Nivolumab ELISA (mAb-based) AB113

REF	IG-AB113	LOT	AB113-105	16.01.2028(dd-mm-yy)	IVD	CE
------------	----------	------------	-----------	----------------------	------------	-----------

Standard Curve				
10X	1:10 Diluted Study Concentration (ng/mL)	Measured Value OD 450/620nm	Range OD	
Standard A	600	3.225	≥ 1.500	
Standard B	200	1.704	-	
Standard C	60	0.645	-	
Standard D	20	0.244	-	
Standard E	0	0.047	≤ 0.100	

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date (dd-mm-yy): 16.01.2026

QC: PASS