



Quality Control Certificate

Antibody to Adalimumab ELISA BB103

REF IG-BB103	LOT BB103-212	 09.01.2028 (dd.mm.yyyy)	IVD 
---------------------	----------------------	---	--

Standards/Positive Control/Cut-Off			
	Ready-to-use Concentration (ng/mL)	Measured Value (OD450nm)	Range OD
Standard A	150	2,295	>0,500
Standard B	50	0,853	>Standard C
Standard C	15	0,328	>Standard D
Standard D	5	0,157	>Standard E
Standard E	0	0,083	<0,100
Calibrator		0.074	<0,100
Positive Control	60,3 ng/mL	0.982	± 20%

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date: 09.01.2026 (dd,mm,yyyy)

QC: PASS

